




Quality Control Results - Table and Graphical View

QC Name: 0409274301
QC Level: 1
QC Lot: 0409274301
Test Name: Glucose

Analyser: Nova StatStrip
Site / Location: Biribi Clinic - Broken Hill

Levy Jennings Graph



<u>Date and Time</u>	<u>Result(s)</u>	<u>Unit</u>	<u>Target Value</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>WestGard</u>	<u>Operator</u>
30 Jun 2010 15:56:05	2.6	mmol/L	3.2	2.6	0	0	ins 1	Community Nurse, PoC
06 Jul 2010 12:10:05	2.6	mmol/L	3.2	2.6	0	0	ins 4	Community Nurse, PoC
16 Jul 2010 08:41:02	2.8	mmol/L	3.2	2.7	0.12	4.33	ins 4	Community Nurse, PoC
29 Jul 2010 13:11:31	2.5	mmol/L	3.2	2.6	0.13	4.79	ins 4	Community Nurse, PoC
05 Aug 2010 11:53:31	2.6	mmol/L	3.2	2.6	0.11	4.18	ins 10	Community Nurse, PoC
12 Aug 2010 12:48:07	2.6	mmol/L	3.2	2.6	0.1	3.76	ins 10	Community Nurse, PoC
20 Aug 2010 11:40:34	2.9	mmol/L	3.2	2.7	0.14	5.26	ins 10	Community Nurse, PoC
31 Aug 2010 12:58:43	3.2	mmol/L	3.2	2.7	0.23	8.49	ins 10	Community Nurse, PoC
03 Sep 2010 13:35:39	2.3	mmol/L	3.2	2.7	0.26	9.66	ins 10	Community Nurse, PoC
08 Sep 2010 14:10:18	3.0	mmol/L	3.2	2.7	0.26	9.76	ins 10	Community Nurse, PoC
08 Sep 2010 14:12:43	3.0	mmol/L	3.2	2.7	0.27	9.71	Passed	Community Nurse, PoC
08 Sep 2010 14:14:25	2.8	mmol/L	3.2	2.7	0.25	9.26	Passed	Community Nurse, PoC
17 Sep 2010 14:01:27	3.0	mmol/L	3.2	2.8	0.25	9.18	Passed	Community Nurse, PoC